



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/530,683	09/02/2005	Frederic Impellizzeri	1811-64	7190

24106 7590 11/01/2007
EGBERT LAW OFFICES
412 MAIN STREET, 7TH FLOOR
HOUSTON, TX 77002

EXAMINER

HOFFMAN, MARY C

ART UNIT	PAPER NUMBER
----------	--------------

3733

MAIL DATE	DELIVERY MODE
-----------	---------------

11/01/2007

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

CT

Office Action Summary

Application No.

10/530,683

Applicant(s)

IMPELLIZZERI, FREDERIC

Examiner

Mary Hoffman

Art Unit

3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 16 August 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 19-23 and 25-28 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 19-23 and 25-28 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 15 April 2007 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some * c) ☐ None of:
1. ☒ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Interview Summary

It is noted that the interview summary mailed 7/20/2007 contained several typographical errors that are herein corrected for the record: In line 4, "non-obvious" should have been stated as --obvious--, and in line 5, "unobvious" should have been stated as --obvious--. The examiner maintains that it was made clear during the interview that the all of the claims were either anticipated or obvious over the prior art despite the typographical errors contained in the interview summary. As a result of the interview, Applicant has significantly amended to try to overcome the cited prior art.

Also, it is further noted that it is the responsibility of the applicant or the attorney or agent to make the substance of an interview of record in the application file. See MPEP 713.04.

Claim Rejections - 35 USC § 112

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claims 27 and 28 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed,

had possession of the claimed invention. The disclosure does not reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the limitations "said plurality of inserts having a generally conical shape" in claim 27, line 3, and "planarly fixed" in claim 28, line 14. The specification does not recite this term, and the drawings do not appear to show generally conical or planarly fixed inserts. Rather, the inserts appear to be rounded and protruding from the planar bone plate surface in FIGS. 4-6. Thus, the terms "generally conical" and "planarly fixed" is being considered new matter. It is further noted that with regard to claim 27, it appears that Applicant may have intended to claim that the inserts have generally conical shaped openings, which would not be considered new matter if recited in the claims.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 19-23 and 25-28 are rejected under 35 U.S.C. 103(a) as being unpatentable over Buhler et al. (U.S. Patent No. 6,540,746) in view of Dixon et al. (U.S. Patent No. 6,656,181) and Frigg et al. (U.S. Patent No. 6,206,881).

Buhler et al. disclose a self-locking osteosynthesis device comprising a plate (ref. #1) having a plurality of openings (ref. #19) formed therein, each of the plurality of

Art Unit: 3733

in the art at the time the invention was made to construct the device of Buhler et al. with the plurality of inserts having a width greater than the plate in view of Dixon et al. to perform the function of guiding.

2.) Frigg et al. disclose tapping bone screw heads having a conical shape and conical threading allowing the bone screws to be angularly received into conical openings to achieve locking even when the bone screw is imperfectly inserted coaxially (see FIG. 1-2, col. 2, lines 1-7 and line 24). It would have been obvious to one having ordinary skill in the art at the time the invention was made to construct the device of Buhler et al. with tapping bone screw heads having a conical shape and conical threading allowing the bone screw to be angularly received into conical openings in view of Frigg et al. to achieve locking even when the bone screw is imperfectly inserted coaxially.

3.) It would have further been obvious to one having ordinary skill in the art at the time the invention was made to construct the device of Buhler et al. in view of Dixon et al. and Frigg et al. with the plurality of inserts being formed of a thermoplastic polymer or polyether ether ketone material, since it has been held to be within the general skill of a worker in the art to select a known material on the basis of its suitability for the intended use as a matter of obvious design choice. In re Leshin, 125 USPQ 416.

Response to Arguments

Applicant's arguments with respect to the claims have been considered but are moot in view of the new ground(s) of rejection.

Art Unit: 3733

in the art at the time the invention was made to construct the device of Buhler et al. with the plurality of inserts having a width greater than the plate in view of Dixon et al. to perform the function of guiding.

2.) Frigg et al. disclose tapping bone screw heads having a conical shape and conical threading allowing the bone screws to be angularly received into conical openings to achieve locking even when the bone screw is imperfectly inserted coaxially (see FIG. 1-2, col. 2, lines 1-7 and line 24). It would have been obvious to one having ordinary skill in the art at the time the invention was made to construct the device of Buhler et al. with tapping bone screw heads having a conical shape and conical threading allowing the bone screw to be angularly received into conical openings in view of Frigg et al. to achieve locking even when the bone screw is imperfectly inserted coaxially.

3.) It would have further been obvious to one having ordinary skill in the art at the time the invention was made to construct the device of Buhler et al. in view of Dixon et al. and Frigg et al. with the plurality of inserts being formed of a thermoplastic polymer or polyether ether ketone material, since it has been held to be within the general skill of a worker in the art to select a known material on the basis of its suitability for the intended use as a matter of obvious design choice. In re Leshin, 125 USPQ 416.

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. See attached PTO-892, particularly Kaute (U.S. Patent No. 3,596,656), Kummer et al. (4,338,926) and Sutter et al. (U.S. Patent No. 4,388,921).

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Mary Hoffman whose telephone number is 571-272-5566. The examiner can normally be reached on Monday-Friday 9:00-5:00pm.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo C. Robert can be reached on 571-272-4719. The fax phone

Art Unit: 3733

number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

MCH



EDUARDO C. ROBERT
SUPERVISORY PATENT EXAMINER